



PRODUCT SPECIFICATION SHEET
WATER

Meets HPLC Grade Monographs
0.22 micron filtered

Main Catalog #: 23200HPLC-Package Size*

Available in the following sizes:

*Refer to Master Price List – Individual package sizes have unique size codes

| Test | Specification | Typical Result |
|---|----------------|----------------|
| Fluorescence at emission maximum for Impurities at 450 nm | 0.1 ppb max. | <0.1 ppb |
| LC Gradient Elution Suitability at 254 nm | To Pass Test | Pass |
| Residue after Evaporation | 1 ppm max. | <0.1 ppm |
| Specific Conductance at 25C ¹ | 1.5 µS/cm max. | 0.05 µS/cm |
| Optical Absorbance at: | | |
| 190 nm | 0.01 max. | 0.00 |
| 200 nm | 0.01 max. | 0.00 |
| 210 nm | 0.01 max. | 0.00 |
| 250 nm | 0.005 max. | 0.001 |
| 400 nm | 0.005 max. | 0.001 |
| Substances reducing Permanganate | To Pass Test | Pass |

Form: Water. HPLC. Rev 1.2, 06/16, KAD

¹ Meets stated value at the time of manufacture.
Not intended for use as a sterile or potable product.

This product is for further commercial manufacturing, laboratory or research use, and may be used as an excipient or a process solvent for pharmaceutical purposes. It is not intended for use as an active ingredient in drug manufacturing nor as a medical device or disinfectant. Appropriate/legal use of this product is the responsibility of the user.