



PRODUCT SPECIFICATION SHEET
PURIFIED WATER

Meets USP Monograph
0.22 micron filtered

Main Catalog #: zp232000USP-Size Code*
Recommended Retest Date: Twelve months from the date of manufacture

Available in the following sizes:

20 Liter Cube

Test	Specification	Typical Result
Appearance	Clear colorless liquid	Pass
pH ¹	5.0 – 7.0	6.0
Total Organic Carbon ¹	<0.50 mg C/L	<0.50 mg C/L
Water Conductivity (stage 2) ¹	NMT 2.1 µS/cm at 25C	1.0 µS/cm
Oxidizable Substances	The pink color does not completely disappear	Passes Test
Microbial Limits	≤10 cfu/mL	≤1 cfu/mL
Residual Solvents	Meets the Requirements	No solvents are used in manufacturing process

Form: Water, USP, Rev 1.0, 10/17, EF

¹ Meets stated value at the time of manufacture.

Not intended for use as a sterile or potable product or in the manufacture of dialysis solutions. Not for Parenteral use, nor for injection, wound irrigation or infusion.

This product is for further commercial manufacturing, laboratory or research use, and may be used as an excipient or a process solvent for pharmaceutical purposes. It is not intended for use as an active ingredient in drug manufacturing nor as a medical device or disinfectant. Appropriate/legal use of this product is the responsibility of the user.